

***National Type Evaluation Program
Certificate of Conformance
for Weighing and Measuring Devices***

For:

Indicating Element
Digital Electronic
Model: HI-3030
 n_{\max} : 10 000

Accuracy Class: III/III L

Submitted by:

Hardy Instruments, Inc.
3860 Calle Fortunada
San Diego, CA 92123-1825
Tel: (858) 278-2900
Fax: (858) 278-9700
Contact: Michael Strauch
e-mail: mstrauch@hardyinst.com

Standard Features and Options

Pound/kilogram unit conversion
Liquid crystal display
Alphanumeric display
Gross/tare/net weight display
Gross/net weight display
Keyboard tare
Semi-automatic (push-button) tare
Semi-automatic (push-button) zero
Automatic zero setting mechanism (AZSM)
AC power supply
Ethernet communication port
Category 3 audit trail

Options:

Remote calibration
Configuration capability

Temperature Range: -10 °C to 40 °C (14 °F to 104 °F)

This device was evaluated under the National Type Evaluation Program (NTEP) and was found to comply with the applicable technical requirements of Handbook 44, "Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.



Don Onwiler
Chairman, NCWM, Inc.



James C. Truex
Chairman, National Type Evaluation Program Committee

Issued Date: November 30, 2005

Note: The National Conference on Weights and Measures does not "approve", "recommend", or "endorse" any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.

Hardy Instruments, Inc
Indicating Element
Model: HI-3030

Application: General purpose indicator for Class III and Class III L installations.

Identification: The identification is on a self-adhesive and self-destructive badge at the bottom of the indicator.

Sealing: Sealing is by Category 3 audit trail only. The configuration and calibration parameters are password protected and can be accessed within the device itself. Sealing is initiated by pressing the “Test or (9)” key and “Enter” key. A “Device Data List” will then be displayed. Press the “Down” arrow key until “View Audit Trail” is displayed. Press “Enter”, the audit trail information is displayed by name, value, and dates in designated audit trail lines. Press “Exit” to leave the audit trail mode.

Operation: The indicator may support up to four weighing elements independently and simultaneously. Each port can be independently calibrated, configured, and sealed for weighing operations.

Test Conditions: The Model HI-3030 electronic indicator was submitted for evaluation. The emphasis of the evaluation was on device design, operation, performance, and compliance with influence factor requirements. The indicator was interfaced with a weight simulator and tested for accuracy over a temperature range of -10 °C to 40 °C (14 °F to 104 °F). Tests were also conducted over a voltage range of 100 VAC to 130 VAC. Additionally, the indicator was interfaced with a weighing element to verify compliance with motion detection, momentary power loss, and zero function requirements.

Type Evaluation Criteria Used: NIST Handbook 44, 2005 Edition, NCWM Publication 14, 2005 Edition

Tested By: S. Boyd (CA)

Conclusion: The results of the evaluations and information provided by the manufacturer indicate the devices comply with applicable requirements.

Information Reviewed By: S. Patoray, L. Bernetich (NCWM)

